

INVITATION FOR QUOTATION

TEQIP-III/WB/fetj/7

17-10-2017

To,

Sub: Invitation for Quotations for Supply of Goods


Package Code: TEQIP-III/2017/fetj/Shopping/8

Dear Sir,

- 1) You are invited to submit your most competitive quotation for the following goods with item wise detailed specifications as given at Annexure - I,

Sr. No	Brief Description	Qty.	Delivery Period (In days)	Place of Delivery	Installation Requirement (if any)
1	Double Beam UV-VIS Spectrophotometer with its software and peripherals	01	45	Department of Food Technology & Bio-Chemical Engineering and	Yes
		01		School of Environmental Studies	

- 2) Government of India has received a credit from the International Development Association (IDA) towards the cost of the Technical Education Quality Improvement Programme [TEQIP]-Phase III Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for quotations is issued.


Nodal Officer (Procurement)
TEQIP Phase-III
Jadavpur University
Kolkata-700 032

3) Quotation:

- a) The contract shall be for the full quantity as described above.
- b) Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
- c) All duties and other levies payable by the supplier under the contract shall be included in the total price.
- d) **Applicable taxes shall be quoted separately for all items.**
- e) The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- f) If you require any "waybill" for transportation of goods to Kolkata, you have to apply in advance separately with a copy of Invoice enclosing a copy of order.
- g) **The Prices should be quoted in Indian Rupees only.**

4) **Each bidder shall submit only one quotation for most suitable one for the item mentioned only, without any optional item.**

5) **Validity of quotation:** Quotation shall remain valid for a period not less than 45 days after the last date of quotation submission.

6) Evaluation of Quotations:


The Purchaser will evaluate and compare the quotations found to be substantially responsive i.e. which

- a) are properly signed ; and
- b) Conforming to the terms and conditions, and specifications.

7) The Quotations would be evaluated for all items together.

- Quotations will be compared on the basis of quoted price (Total Cost including all taxes) for goods at its final destination.
- Past performance & experience may be furnished to consider the credential of the bidder

8) Award of contract: The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.


Nodal Officer (Procurement)
TEQIP Phase-III
Jadavpur University
Kolkata-700 032

a) Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.

b) The bidder whose bid is accepted will be notified for the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.

9) **Delivery:**

9.1 Delivery period mentioned in the Purchase Order received by the vendor must be maintained. In case of failing the delivery within stipulated date the order maybe canceled if proper justification is not received from the vendor in time.

10) **Payment:**

10.1) Payment shall be made in Indian Rupees as follows:

100% of the total cost after delivery, satisfactory installation along with demonstration and testing of the goods and also acceptance by the department.

10.2) Payment will be done by online transfer to the bank account of the vendor by NPIU, New Delhi after formalities. Vendor must provide the following details for bank transfer. (a) Account holder Name, (b) Account No., (c) Bank Name, (d) Branch Name, (e) IFSC No., (f) PAN, (g) TAN, (h) GSTIN, (i) TIN, (j) Mobile number.

11) All supplied items are under normal commercial warranty/ guarantee of not less than 12 months from the date of successful acceptance of items, unless specified of warranty for longer duration in the specification.

12) Rate of AMC may be furnished separately, which will be applicable at the end of the normal commercial warrant period.

13) You are requested to submit your offer on or before 15:30 hours on 02-11-2017. Bids will be opened on the same day at 16:00 hours.

14) Detailed specifications of the items are at Annexure I.

15) Training Clause (if any) **Demonstration at the time of Installation are to be confirmed.**

16) Testing/Installation Clause (if any) **Satisfactory Installation & Testing is required**



Nodal Officer (Procurement)
TEQIP Phase-III
Jadavpur University
Kolkata-700 032


17) Submission of video footage of the successful installation and testing along with Model No. & SL. No. is preferred.

18) Information brochures/ Product catalogue, if any must be accompanied with the quotation clearly indicating the model quoted for.

19) Sealed quotation to be submitted/ delivered, within 15:30 hours of 02-11-2017, at the address mentioned below,

TEQIP CELL, 2nd Floor, Aurobindo Bhavan, Jadavpur University, 188, Raja S. C. Mallick Road, Kolkata- 700 032.

We look forward to receive your quotation and thank you for your interest in this project.



Prof. Sanjib Acharyya

Nodal Officer, Procurement

TEQIP, Phase-III, JU

Nodal Officer (Procurement)
TEQIP Phase-III
Jadavpur University
Kolkata-700 032



Annexure I

Package Code: TEQIP-III/WB/fetj/7

Sr. No	Item Name	Specifications																												
1	Double Beam UV-VIS Spectrophotometer with its software and peripherals	<table border="1"> <thead> <tr> <th colspan="2" data-bbox="596 528 1490 562">UV-VIS DUAL BEAM SPECTROPHOTOMETER SPECIFICATIONS:</th> </tr> </thead> <tbody> <tr> <td data-bbox="596 562 932 595">Optical Design</td> <td data-bbox="932 562 1490 595">Dual beam—internal reference detector</td> </tr> <tr> <td data-bbox="596 595 932 629">Spectral Bandwidth</td> <td data-bbox="932 595 1490 629">1.8 nm</td> </tr> <tr> <td data-bbox="596 629 932 685">Light Source (Typical Lifetime)</td> <td data-bbox="932 629 1490 685">Xenon flash lamp (5 years)</td> </tr> <tr> <td data-bbox="596 685 932 719">Detector</td> <td data-bbox="932 685 1490 719">Dual silicon photodiodes</td> </tr> <tr> <td data-bbox="596 719 932 842">Wavelength Accuracy Slew Speed Data Interval</td> <td data-bbox="932 719 1490 842">Range Repeatability Scanning 190 to 1100 nm ± 1.0 nm ± 0.5 nm 11,000 nm/min 10 to 4,200 nm/min 0.2, 0.5, 1.0, 2.0, 3.0, 5.0 nm</td> </tr> <tr> <td data-bbox="596 842 932 987">Photometric Measurement Modes Range Linearity Accuracy Noise Drift</td> <td data-bbox="932 842 1490 987">Absorbance, % transmittance, concentration -0.5 to 5.0 A ; -1.5 to 125 %T ; ± 9999 C Up to 3.5 A at 260 nm ± 0.005 A at 1.0 A, <0.00025 at 0.0 A <0.00050 at 1.0 A, <0.00080 at 2.0 A RMS at 260 nm <0.0005 A/hr.</td> </tr> <tr> <td data-bbox="596 987 932 1055">Stray Light</td> <td data-bbox="932 987 1490 1055"><0.08 %T at 220 nm and 340 nm, <0.1 %T at 340 nm and 400 nm, <1.0 %T at 198 nm</td> </tr> <tr> <td data-bbox="596 1055 932 1099">Display</td> <td data-bbox="932 1055 1490 1099">Graphical with LCD backlight</td> </tr> <tr> <td data-bbox="596 1099 932 1155">Keypad</td> <td data-bbox="932 1099 1490 1155">Sealed membrane with tactile response keys</td> </tr> <tr> <td data-bbox="596 1155 932 1256">Connectivity</td> <td data-bbox="932 1155 1490 1256">USB type A port for USB stick (front panel), USB type B port for computer (rear panel), USB type A port for printer (rear panel)</td> </tr> <tr> <td data-bbox="596 1256 932 1301">Power Requirements</td> <td data-bbox="932 1256 1490 1301">100 to 240 V ; 50 to 60 Hz</td> </tr> <tr> <td data-bbox="596 1301 932 1346">Warranty</td> <td data-bbox="932 1301 1490 1346">1 year</td> </tr> <tr> <td data-bbox="596 1346 932 1536">Supply should be included</td> <td data-bbox="932 1346 1490 1536">UV-VIS DUAL BEAM Spectrophotometer, includes methods on USB memory stick, 6-position cell holder, 3-position cell holder, 1-position cell holder, user documentation, dust cover, USB cable and 110 V and 220 V power cords</td> </tr> </tbody> </table>	UV-VIS DUAL BEAM SPECTROPHOTOMETER SPECIFICATIONS:		Optical Design	Dual beam—internal reference detector	Spectral Bandwidth	1.8 nm	Light Source (Typical Lifetime)	Xenon flash lamp (5 years)	Detector	Dual silicon photodiodes	Wavelength Accuracy Slew Speed Data Interval	Range Repeatability Scanning 190 to 1100 nm ± 1.0 nm ± 0.5 nm 11,000 nm/min 10 to 4,200 nm/min 0.2, 0.5, 1.0, 2.0, 3.0, 5.0 nm	Photometric Measurement Modes Range Linearity Accuracy Noise Drift	Absorbance, % transmittance, concentration -0.5 to 5.0 A ; -1.5 to 125 %T ; ± 9999 C Up to 3.5 A at 260 nm ± 0.005 A at 1.0 A, <0.00025 at 0.0 A <0.00050 at 1.0 A, <0.00080 at 2.0 A RMS at 260 nm <0.0005 A/hr.	Stray Light	<0.08 %T at 220 nm and 340 nm, <0.1 %T at 340 nm and 400 nm, <1.0 %T at 198 nm	Display	Graphical with LCD backlight	Keypad	Sealed membrane with tactile response keys	Connectivity	USB type A port for USB stick (front panel), USB type B port for computer (rear panel), USB type A port for printer (rear panel)	Power Requirements	100 to 240 V ; 50 to 60 Hz	Warranty	1 year	Supply should be included	UV-VIS DUAL BEAM Spectrophotometer, includes methods on USB memory stick, 6-position cell holder, 3-position cell holder, 1-position cell holder, user documentation, dust cover, USB cable and 110 V and 220 V power cords
UV-VIS DUAL BEAM SPECTROPHOTOMETER SPECIFICATIONS:																														
Optical Design	Dual beam—internal reference detector																													
Spectral Bandwidth	1.8 nm																													
Light Source (Typical Lifetime)	Xenon flash lamp (5 years)																													
Detector	Dual silicon photodiodes																													
Wavelength Accuracy Slew Speed Data Interval	Range Repeatability Scanning 190 to 1100 nm ± 1.0 nm ± 0.5 nm 11,000 nm/min 10 to 4,200 nm/min 0.2, 0.5, 1.0, 2.0, 3.0, 5.0 nm																													
Photometric Measurement Modes Range Linearity Accuracy Noise Drift	Absorbance, % transmittance, concentration -0.5 to 5.0 A ; -1.5 to 125 %T ; ± 9999 C Up to 3.5 A at 260 nm ± 0.005 A at 1.0 A, <0.00025 at 0.0 A <0.00050 at 1.0 A, <0.00080 at 2.0 A RMS at 260 nm <0.0005 A/hr.																													
Stray Light	<0.08 %T at 220 nm and 340 nm, <0.1 %T at 340 nm and 400 nm, <1.0 %T at 198 nm																													
Display	Graphical with LCD backlight																													
Keypad	Sealed membrane with tactile response keys																													
Connectivity	USB type A port for USB stick (front panel), USB type B port for computer (rear panel), USB type A port for printer (rear panel)																													
Power Requirements	100 to 240 V ; 50 to 60 Hz																													
Warranty	1 year																													
Supply should be included	UV-VIS DUAL BEAM Spectrophotometer, includes methods on USB memory stick, 6-position cell holder, 3-position cell holder, 1-position cell holder, user documentation, dust cover, USB cable and 110 V and 220 V power cords																													

Nodal Officer (Procurement)
TEQIP Phase-III
Jadavpur University
Kolkata-700 032



FORMAT FOR QUOTATION SUBMISSION

(In letterhead of the supplier with seal)

Package Code:TEQIP-III/WB/fetj/7

Date: _____

To: _____

Sl. No.	Description of goods (with full Specifications)	Qty.	Unit	Quoted Unit rate in Rs. (Including Ex Factory price, excise duty, packing and forwarding, transportation, insurance, other local costs incidental to delivery and warranty/ guaranty commitments)	Total Price (A)	Applicable Taxes	
						In %	In figures (B)
Total Cost							

Gross Total Cost (A+B): Rs. _____

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs. -----
(Amount in figures) (Rupees ----- amount in words) within the period specified in the Invitation for Quotations.

We confirm that the normal commercial warranty/ guarantee of ----- months shall apply to the offered items and we also confirm to agree with terms and conditions as mentioned in the Invitation Letter.

We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery.

Signature of Supplier

Name: _____

Address: _____

Contact No: _____